

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

DEC 21 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130189
Name	Richard Eaton	ID#	7287
Date	12/11/09		
Agency	Saint Paul Island DB	Phone #	907-546-3130
Instrument Location	Police Dept		
Alco S/N		Target Value	High Pressure
56801		.081	200
Alco Test Values		1 st Alco	2 nd Alco
		.078	.078
Signature		Richard B Eaton	
(OVER)			

12-22-09
Cm

BMB
12/31/09

(Do Not write in the area below)


- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31st day of Dec, 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130189

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY

INSTRUMENT LOG NUMBER

NOVEMBER 30, 2009

OPERATOR'S NAME:

EATON RICHARD BURTON

MEDICAL'S MINOR: 001

DATE OF THE LAST TEST:

ELECTRONIC FIRST NAME ID:

TEST #1: 00000

TEST #2: 00001

TEST #3: 00002

TEST #4: 00003

TEST #5: 00004

TEST #6: 00005

TEST #7: 00006

--- BREATH ANALYSIS ---

TEST #1: 00000

TEST #2: 00001

TEST #3: 00002

TEST #4: 00003

TEST #5: 00004

TEST #6: 00005

TEST #7: 00006

TEST #8: 00007

TEST #9: 00008

TEST #10: 00009

TEST #11: 00010

TEST #12: 00011

TEST #13: 00012

TEST #14: 00013

TEST #15: 00014

TEST #16: 00015

TEST #17: 00016

TEST #18: 00017

TEST #19: 00018

TEST #20: 00019

TEST #21: 00020

TEST #22: 00021

TEST #23: 00022

TEST #24: 00023

TEST #25: 00024

TEST #26: 00025

TEST #27: 00026

TEST #28: 00027

TEST #29: 00028

TEST #30: 00029

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY

INSTRUMENT LOG NUMBER

NOVEMBER 30, 2009

TIME 00:00

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS: OKAY

SAMPLE CHAMBER: 40c

BREATH TUBE: 41c

BAROMETER: 29.15 in

AIR DETECTOR: OKAY

PUMP: OKAY

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

11/30/09 00:00:00 130189 00000

11/30/09 00:00:00 130189 00001

11/30/09 00:00:00 130189 00002

11/30/09 00:00:00 130189 00003

11/30/09 00:00:00 130189 00004

11/30/09 00:00:00 130189 00005

11/30/09 00:00:00 130189 00006

11/30/09 00:00:00 130189 00007

11/30/09 00:00:00 130189 00008

11/30/09 00:00:00 130189 00009

11/30/09 00:00:00 130189 00010

11/30/09 00:00:00 130189 00011

11/30/09 00:00:00 130189 00012

11/30/09 00:00:00 130189 00013

11/30/09 00:00:00 130189 00014